# SuperTran VP-cryostats STVP-NMR







# SuperTran VP-cryostats STVP-NMR




## **Specifications**

	STVP-NMR
Temperature range	<2 K (1.5 K in single-shot mode) to 325 K (420 K optional)
Typical temperature stability <sup>1</sup>	±50 mK
Orientation <sup>2</sup>	Vertical for operation <4.5 K
Cooldown time (LHe to 5 K)	30 min
Cryogen consumption (LHe room temp to 4.2 K)	0.5 L
Cryogen consumption (LHe at 5 K)	1.4 L/h
Height (approximate)	Customer-specified to fit magnet
Inner space (at sample region)	55 mm (2.16 in)
Weight (excluding transfer line, approximate)	7 kg (15.4 lb)
Shipping weight (cryostat + line, approximate)	79 kg (174 lb)
Shipping dimensions (cryostat + line, approximate)	$1905 \times 990.6 \times 431.8$ mm (75 $\times$ 39 $\times$ 17 in)

<sup>&</sup>lt;sup>1</sup> Measured with temperature controller

<sup>&</sup>lt;sup>2</sup> Cryogen consumption may be higher during non-vertical operation





## SuperTran VP-cryostats STVP-NMR

### Complete your setup

#### Temperature control

Included



Every cryostat includes a Lake Shore temperature controller and calibrated sensor.

#### MeasureLINK control software


Optional add-on



MeasureLINK software enables a wide range of capabilities including charting and logging, system monitoring with a cryostat-specific process view, and controlling Lake Shore equipment as well as third-party instrumentation. No programming required—drag-and-drop to create temperature sweeps, access measurements, and see real-time internal cryostat temperatures in process view.

#### Cryogen-free operation

Optional add-on



Cryostats can be combined with the RGC recirculating gas cooler for fully cryogen-free operation throughout the entire temperature range. This enables unattended cryostat operation, ideal for extended duration measurements.



## SuperTran VP-cryostats STVP-NMR



## Configure your cryostat

#### 1. Select cryostat variant

STVP-NMR Optical, <2 K to 325 K, calibrated Cernox®

Custom configurations are available to fit your experiment needs — contact Sales for details

#### 2. Select cryostat configurations

#### **Optional bottom window**

See our cryostat window selection guide for additional information.

WR-STD-FS Fused silica
WR-UV-FS UV-grade fused silica

WR-STD-SAPH Sapphire WR-STD-ZNSE ZnSe

#### 3. Select pump (optional)

Each cryostat requires a pump to operate. If you do not have an existing pump to use, select one of the pumps below.

10RVP General-purpose mechanical pumping station
10DDP General-purpose mechanical pumping station

with LN<sub>2</sub> cold trap and isolation valve

TS-85-D Turbopumping station

#### 4. Select cryostat wiring

We offer a variety of both unwired and wired feedthroughs to complete your measurement setup. Please refer to the cryostat feedthroughs and wiring guide for more information.

#### 5. Select optional setup configurations

#### Cryogen-free operation

RGC4-10 Recirculating cooler with base temperature

<10 K

RGC4-15 Recirculating cooler with base temperature <8 K
RGC4-20 Recirculating cooler with base temperature <7 K

#### Measurement instrumentation

Cryostats come standard with one temperature controller.

Model 336 temperature controller
 Model 335 temperature controller
 Model 325 temperature controller

#### 6. Select optional control software

ML-MCS MeasureLINK-MCS software with scripting

development license; includes lifetime activation for version purchased and full MeasureLINK capability on up to 5 computers with Lake Shore instrument drivers, chart recorder functionality, and drag-and-drop measurement sequences; some application packs sold separately

#### 7. Select additional accessories

Cryostats come standard with one installed temperature sensor. Other sensors are available—contact us.

**CX-1050-CU-HT-1.4M** Cernox® magnetic field independent, calibrated

CF-100 LHe storage Dewar

LN-50 LN<sub>2</sub> storage Dewar configured for use with

SuperTran cryostats

Copyright © Lake Shore Cryotronics, Inc. All rights reserved. Specifications are subject to change.

102424 10:27





